



## NATIONAL TYPE EVALUATION PROGRAM

# Certificate of Conformance

for Weighing and Measuring Devices

**For:**

Load Cell  
Single-Ended Shear Beam, Compression  
Model: CG-23, CG-SB, CG-MK15, CG-SSB, CG-743 & CG-745 (formerly SE-CG Series)  
nmax: Multiple Cell 5 000  
vmin: (see page 2)  
Capacity: 250 lb to 20 000 lb  
Accuracy Class: III

**Submitted By: Contact Info. Updated Oct. 2014**

Coti Global Sensors, Inc.  
5699 Highway 53  
Harvest, AL 35773  
Tel: 256-852-9900  
Fax: 256-852-9903  
Contact: Amy Allen  
Email: [amy@cotiglobal.com](mailto:amy@cotiglobal.com)  
Web site: [www.cotiglobal.com](http://www.cotiglobal.com)

**Standard Features and Options****Standard Features:**

- Alloy Steel or Stainless Steel Construction
  - Method of Sealing: Potted with Metal Cover
  - Number of Wires: 4 wires
  - Excitation Voltage: 10 VDC
  - Nominal Output: 3.0 mV/V or 2.0 mV/V
- Bridge Resistance Input Nominal: 350 Ohms

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. \*Editorial changes, not affecting the type or metrological content, corrected this certificate.

John Gaccione  
Chairman, NCWM, Inc.

Stephen Benjamin  
Committee Chair, National Type Evaluation Program Committee

Issued: April 17, 2014

**1135 M Street, Suite 110 / Lincoln, Nebraska 68508**

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



**Coti Global Sensors, Inc.**

Load Cell / CG-23, CG-SB, CG-MK15, CG-SSB, CG-743 & CG-745 Series

**Application:** The load cells may be used in Class III scales for multiple cell applications consistent with the model designations, number of scale divisions, and parameters specified in this Certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the  $v_{min}$  values, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions ( $n_{max}$ ) and with larger  $v_{min}$  values than those listed on the Certificate. However, the load cells must be marked with the appropriate  $n_{max}$  and  $v_{min}$  for which the load cell may be used.

**Identification:** A pressure sensitive, tamper evident, identification badge containing the manufacturer, model designation, and serial number is located on the load cell. All other required information, if not marked on the load cell, must be on an accompanying document including the serial number of the load cell.

Models	Capacity (lb)	$v_{min}$ (lb)	Minimum Dead Load (lb)
CG-23	250	0.038	10
CG-23-SS	500*	0.075	10
CG-23-LP	1000	0.14	10
CG-23-SSW	1250	0.175	10
CG-SB3	1500	0.21	10
CG-MK15	2000	0.28	10
CG-SB250	2500	0.35	50
CG-SSB	3000	0.42	50
CG-743	4000	0.56	88
CG-745	5000*	0.70	88
	10 000	1.40	88
	15 000	2.10	88
	20 000	2.80	88

\* Two Load Cells Tested

**Test Conditions:** This certificate supersedes Certificate of Conformance Number 08-074 and was issued to recognize a change to the model number designations by removing the SE prefix. No additional testing was deemed necessary. Contact information was also updated. Previous test conditions are listed below for reference.

**Certificate of Conformance Number 08-074:** Two 500 lb capacity and two 5000 lb load cells were tested at NIST using dead weights as the reference standard. The data were analyzed for multiple load cell applications. The cells were tested over a temperature range of -10 °C to 40 °C. Three tests were run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test was waived due to the insensitivity of the load cell design to changes in barometric pressure.

**Evaluated By:** NIST Force Group, NIST Office of Weights and Measures 08-074

**Type Evaluation Criteria Used:** NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2014. NCWM, Publication 14: Weighing Devices, 2014.

**Conclusion:** The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

**Information Reviewed By:** S. Patoray, L. Bernetich (NCWM) 04-066; J. Truex (NCWM) 08-074, 08-074A1



**Coti Global Sensors, Inc.**

Load Cell / CG-23, CG-SB, CG-MK15, CG-SSB, CG-743 & CG-745 Series

**Example of Device:**

